

**Product Name :**  
Flexural Test Devices for Mortar Beams

**Product Code :**  
ENGLABINGCAG3900025



**Description :**

Flexural Test Devices for Mortar Beams

**Technical Specification :**

One of the two lower bearer can tilt horizontally and the distance between the two bearers is 100 mm and 119 mm. Apart from that, the two models are identical.

Robust frame fitted with an upper bearer that moves vertically supported by springs.

Total height is 188 mm

Weight approx.: 8 kg